

DECLARATION of CONFORMITY

We, NIKON INSTRUMENTS EUROPE B.V.
LAAN VAN KRONENBURG 2, 1183AS, AMSTELVEEN,
THE NETHERLANDS ,

declare under our sole responsibility that the product,

TYPE of PRODUCT: Microscope

MODEL NUMBER: ECLIPSE Ci-L

to which this declaration relates is in conformity
with the following standards or other normative documents,

EN 61010-1:2001, EN 61010-2-101:2002, EN 591:2001

EN 62366:2008 EN 62311 :2008

EN 61326-1:2006, EN 61326-2-6:2006

EN 55011:2009+A1:2010 Group 1 / Class B

EN 61000-4-2:2009

EN 61000-4-3:2006+A1:2008+A2:2010

EN 61000-4-4:2004+A1:2010

EN 61000-4-5:2006 EN 61000-4-6:2009

EN 61000-4-8:2010 EN 61000-4-11:2004

EN 61000-3-2:2006+A1:2009+A2:2009 EN 61000-3-3:2008

following the provisions of

IN VITRO DIAGNOSTIC MEDICAL DEVICE Directive (98/79/EC)

LOW VOLTAGE Directive (2006/95/EC)

EMC Directive (2004/108/EC)

Manufacturer's name: NIKON CORPORATION
INSTRUMENTS COMPANY

Location of
the manufacturer: 471 NAGAODAI-MACHI,
SAKAE-KU, YOKOHAMA-SHI,
KANAGAWA 244-8533 , JAPAN

AMSTELVEEN /
August 5, 2011


S. Eimori, President