



Leica DM750 User Manual

ryf ag
ryf
Ryf AG
Bettlachstrasse 2
2540 Grenchen
tel 032 654 21 00
fax 032 654 21 09
www.ryfag.ch

Leica
MICROSYSTEMS

Manufacturer Information

Published in January 2015 by:

*Leica Microsystems (Schweiz) AG
Max Schmidheiny Strasse 201
CH-9435 Heerbrugg (Switzerland)*

In case of questions, please contact the hotline:
sms.support@leica-microsystems.com

Chapter Overview

Safety Regulations	5
The Leica DM750	16
Ready!	19
Set!	29
Go!	42
Care of the Microscope	44
Troubleshooting	47
Dimensions	50

Contents

Safety Regulations	
Safety Concept	6
Symbols Used in This Instruction Manual	7
Important Notes	8
Instructions on Use	9
Health Risks and Dangers of Use	11
Information for the Person Responsible for the Instrument	12
Care Instructions	13
Accessories, Maintenance and Repair	14
Electrical Data and Ambient Conditions	15

The Leica DM750	
Introduction	17
Unpacking	18

Ready!	
Substage Illumination	20
Attaching the Viewing Tubes	21
Leica EZ Viewing Tube – Integrated Eyepieces	22
Leica Standard Viewing Tube – Separated Eyepieces	24
Eyecups	25
Installing Objectives	26
Installing the Substage Condenser	27

Set!	
Turning on the Microscope	30
Complete Condenser Centering	31
Using the Condenser	32
Prepare to View a Specimen Slide	33
Focusing	34
Viewing Tube Adjustment	35
Koehler Configuration	37
Oil Immersion Technique	39
Time Delay Shutoff	41

Go!	
Ready! Set! Go!	43

Care of the Microscope	
General Maintenance	45

Troubleshooting	
Troubleshooting	48

Dimensions	
Dimensions	51

Safety Regulations

Safety Concept

The individual modules of the Leica DM microscopy series include an interactive CD-ROM with all relevant user manuals in several languages. Keep it in a safe place, and readily accessible to the user. User manuals and updates are also available for you to download and print from our website at www.leica-microsystems.com.

This user manual describes the special functions of the individual modules of the Leica DM microscopy series and contains important instructions for their operational safety, maintenance, and accessories.

The "Safety concept" booklet contains additional safety information regarding the service work, requirements and the handling of microscope, accessories and electrical accessories as well as general safety instructions. You can combine individual system articles with articles from external suppliers. Please read the user manual and the safety requirements of the supplier.

Before installing, operating or using the instruments, read the user manuals listed above. In particular, please observe all safety instructions.

To maintain the unit in its original condition and to ensure safe operation, the user must follow the instructions and warnings contained in these user manuals.

Symbols Used in This Instruction Manual

Warning of a danger



This symbol indicates especially important information that is mandatory to read and observe.

Failure to comply can cause the following:

- Hazards to personnel
- Functional disturbances or damaged instruments

Warning of hazardous electrical voltage



This symbol indicates especially important information that is mandatory to read and observe.

Failure to comply can cause the following:

- Hazards to personnel
- Functional disturbances or damaged instruments

Danger due to hot surface



This symbol warns against touching accessible hot surfaces, e.g. those of light bulbs.

Important information



This symbol indicates additional information or explanations that intend to provide clarity.

Explanatory notes

- ▶ This symbol within the text points to additional information and explanations.



Instructions for disposing of the instrument, accessory components, and consumables.

Important Notes

Description

The Leica DM750 microscope meets today's state of the art of technology. Nevertheless, hazards may still arise during operation. The potential risks are described below.



Before installing, operating or using the instrument, it is mandatory to read this user manual. In particular, please observe all safety instructions.

User manual

This user manual includes important instructions related to operating safety, maintenance and accessories.



Your Leica DM750 microscope comes with an interactive CD-ROM with all relevant user manuals. Keep it in a safe place, and readily accessible to the user. User manuals and updates are also available for you to download and print from our website at www.leica-microsystems.com.

Accessories from third-party suppliers

You can combine individual system articles with articles from external suppliers. Please read the user manual and the safety requirements of the supplier.

Original condition

To maintain the unit in its original condition and to ensure safe operation, the user must follow the instructions and warnings contained in these user manuals.

Legal requirements

Adhere to general and local regulations relating to accident prevention and environmental protection.

EC Declaration of Conformity

Electrically operated accessories are constructed based on the state of the art of technology and are provided with an EC Declaration of Conformity.



Instructions on Use



The Leica DM750 microscope may be used only in closed rooms and must be placed on a solid substrate.

The Leica DM750 microscope can be used in clean rooms without any problems.



Always position the Leica DM750 microscope so that you can disconnect it from the power supply at any time. The power cable must remain accessible at all times, because it is intended as a power disconnect device.

Place of use

Only use the instruments in closed, dust free rooms and between +10°C and +40°C. Protect the devices from oil, chemicals and extreme humidity. If using the devices outdoors, protect them from dust and moisture. Never use electrical devices outdoors. Install electrical devices at least 10 cm from the wall and away from flammable substances.

Avoid large temperature fluctuations, direct sunlight and vibrations. These conditions can distort measurements and micrographic images.



In warm and warm-damp climatic zones, the individual components require special care in order to prevent the build-up of fungus.

Non-intended use



Never install any other plug or unscrew any mechanical components unless expressly instructed to do so in the instructions.



The devices and accessories described in this instruction manual have been tested for safety and potential hazards.



The responsible Leica affiliate must be consulted whenever the instrument is altered, modified or used in conjunction with non-Leica components that are outside of the scope of this manual!



Unauthorized alterations to the instrument or noncompliant use shall void all rights to any warranty claims.



Instructions on Use (Continued)

Transport

If at all possible, use the original packaging for shipping or transporting individual modules.

In order to prevent damage from vibrations, disassemble all moving parts that (according to the user manual) can be assembled and disassembled by the customer and pack them separately.

Disposal

Once the product has reached the end of its service life, please contact Leica Service or Sales about disposal.

Please observe and ensure compliance with the national laws and regulations that implement, for example, the EC Directive WEEE.



Like all electronic devices, this instrument, its accessory components and consumables must never be disposed of with general household waste. Disposal must comply with locally applicable laws and regulations.

Integration in third-party products

When installing Leica products into third-party products, the manufacturer of the complete system or its dealer is responsible for following all applicable safety instructions, laws and guidelines.



Health Risks and Dangers of Use

Health risks



Workplaces with microscopes facilitate and improve the viewing task, but they also impose high demands on the eyes and holding muscles of the user. Depending on the duration of uninterrupted work, asthenopia and musculoskeletal problems may occur. For this reason, appropriate measures for reduction of the workload must be taken:

- Optimum workplace layout
- Frequent changes of activity
- Thorough training of the personnel, giving consideration to ergonomic and organizational aspects

The ergonomic design and construction of the Leica microscopy series are intended to reduce the exertion of the user to a minimum.

Danger of infection



Direct contact with eyepieces is a potential transmission method for bacterial and viral infections of the eye.



The risk can be kept to a minimum by using personal eyepieces for each individual or detachable eyecups.

Dangers during use

- The Leica DM750 microscope may only be connected to a grounded socket.
- The Leica DM750 microscope may not be operated unless it is in proper functioning condition.

The microscope illumination is in the exempt group (risk group 0) according to EN 62471:2008 when used according to its intended use.



Never look directly into the LED beam of the illumination equipment – either with or without optical instruments – as this increases the risk class. Failure to observe this notice poses a risk of eye damage.



Information for the Person Responsible for the Instrument

Information for the person responsible for the instrument

- Ensure that the Leica DM750 microscope is used only by qualified personnel.
- Ensure that this user manual is always available at the Leica DM750 microscope.
- Carry out regular inspections to make certain that the authorized users are adhering to safety requirements.
- When instructing new users, do so thoroughly and explain the meanings of the warning signs and messages.
- Assign individual responsibilities for starting, operating and servicing the instrument and monitor the observance of these responsibilities.
- Do not use the Leica DM750 microscope unless it is in perfect condition.
- Inform your Leica representative or Leica Microsystems (Schweiz) AG, 9435 Heerbrugg, Switzerland, immediately of any product defect that could potentially cause injury or harm.
- If you use accessories made by third-party manufacturers with the Leica DM750 microscope, be sure that each such manufacturer confirms the safety-engineering, harmless usability of the product and observe the product's user manual.
- Modifications and maintenance of the Leica DM750 microscope may only be performed by professionals expressly authorized by Leica.
- Only original Leica replacement parts may be used in servicing the product.
- After service work or technical modifications, the unit must be reconfigured with observance to our technical requirements.
- If the unit is modified or serviced by unauthorized persons, is improperly maintained (as long as maintenance was not carried out by us), or is handled improperly, Leica will not accept any liability.
- The electric installation in the building must conform to the national standard, e.g. current-operated ground leakage protection (fault-current protection) is suggested.



Care Instructions

General instructions

- Protect the Leica DM750 microscope against damp, vapors, acids, alkalis, and corrosive substances. Do not store chemicals in the vicinity.
- Protect the Leica DM750 microscope from oil and grease. Never grease or oil mechanical parts or sliding surfaces.
- Follow the instructions of the disinfectant manufacturer.
- It is advisable to enter a service agreement with Leica Service.

Cleaning coated parts and plastic parts

- Dust and dirt particles should be removed with a soft brush or lint-free cotton cloth.
- Remove coarse debris with a moistened disposable cloth.
- Acetone, xylene or nitro-containing thinners must NOT be used.
- Never use chemicals to clean colored surfaces or accessories with rubberized parts. This could damage the surfaces, and specimens could be contaminated by abraded particles.

Cleaning glass surfaces

- Remove dust using a dry and grease-free brush made from hair, by blowing with bellows, or by using a vacuum.
- Optical surfaces should be cleaned with a lint-free cloth, lens tissue, or cotton swab moistened with a commercially available glass cleaner.



Accessories, Maintenance and Repair

Accessories

Only the following accessories may be used with the Leica DM750 microscope:

- The Leica accessories described in this user manual.
- Other accessories, provided that these have been expressly approved by Leica as being technically safe in this context.

Maintenance

- The Leica DM750 microscope is basically maintenance-free. To ensure that it always operates safely and reliably, we recommend that you take the precaution of contacting the responsible service organization.



You can arrange for periodic inspections or, if appropriate, conclude a maintenance contract with them.

- It is advisable to enter a service agreement with Leica Service.
- For maintenance and repair, only OEM spare parts may be used.

Repairs and service work

- Only original Leica Microsystems spare parts may be used.
- Before opening the instruments, switch off the power and unplug the power cable.
- Avoid contact with powered electrical circuits, which can lead to injury.

Service address

In case of problems, please contact us as follows:

stereo.service@leica-microsystems.com



Electrical Data and Ambient Conditions

Fuse replacement



Unplug the instrument before changing any fuses. The Leica DM750 has two fuses, which are located behind the power cord receptacle.



Only use the following fuse types: 5×20 mm, 1 A/250 V, fast-acting fuse (# 13RFAG30003)

Electrical data

Input: 100–240 V, 50/60 Hz, 5 W (3 W LED)

General safety notes

This instrument of safety class 1 has been built and tested in accordance with the following safety requirements for electrical equipment for measurement, control, and laboratory use:

EN 61010-1: 2002-08	EN 55011: 2007+A2: 2010-05
EN 61010-2-101: 2008-06	EN 60825-1: 2008-05
IEC 61010-1: 2010-06	IEC 60825-1: 2007-03
EN 61326-1: 2006-10	LED Class 1
EN 61326-2-6: 2006-10	



In order to maintain this condition and to ensure safe operation, the user must follow the instructions and warnings contained in this instruction manual.

Environment

Temperature for use	+10 °C ... +40 °C
Storage temperature	-20 °C ... +52 °C +50 °F to +104 °F
Manipulation shock	25 mm on 50 mm hard wood
Transport shock (unpacked)	100 g / 6 ms
Transport shock (packed)	800 mm free fall
Transport vibrations (unpacked)	5–200 Hz / 1.5 g
Air pressure during use and storage	500–1,060 mbar
Humidity during use and storage	20–90 %
Installation Category II (Overvoltage Category)	
Pollution degree 2	



The Leica DM750

Introduction

Thank you for purchasing the Leica DM750 Compound Microscope from Leica Microsystems. This model's exclusive design features and full range of accessories make it a truly versatile, high quality-instrument.



The instrument is to be used only as described. Hazards to personnel may result if used improperly.

Unpacking


- Carefully remove the microscope and any components from the packing carton.
- Verify that all components are intact.
- Check the components against the planned configuration.
- Optional items such as contrast accessories, camera adapters, cameras, and carrying cases are not shipped as part of the standard equipment. These items are delivered in separate packages.
- Please do not discard any of the packing materials. They should be used for safely storing and transporting the instrument should the need arise.



Ready!

Substage Illumination

Substage illumination

 The Leica DM750 is available with two types of Substage Illumination. Identify which type of Illumination you have, as this will be important to know later.

Type 1: Standard illumination

Adjustable condenser centering with provided tool.



Type 2: Koehler illumination

Adjustable condenser centering with thumb-screws and adjustable Koehler field diaphragm.




Attaching the Viewing Tubes

Tools used

- Allen key



 There are two types of viewing tube. Identify which viewing tube you have:



Type 1: Leica EZ viewing tube with integrated eyepieces



Type 2: Standard viewing tube with separate eyepieces

1. Loosen the setscrew (on top of the stand) using the Allen key provided.




2. Set the dovetail into the tripod mount and tighten the setscrew carefully (without using excessive force). This precisely positions the viewing tube onto the optical axis of the microscope, regardless of the selected rotation.



Leica EZ Viewing Tube – Integrated Eyepieces





Leica EZ viewing tube with integrated eyepieces

 To rotate the Leica EZ viewing tube, either loosen the set screw on the stand or replace the set screw with the (optional) captive thumbscrew.

1. In order to use the captive wingscrew, remove the setscrew that was delivered with the stand.



 Make sure that the wingscrew part of the captive wingscrew has been completely separated before you use the stand:




2. Reinstall the Leica EZ viewing tube onto the stand.




3. Tighten the captive wingscrew using the wrench included in the delivery package.



Leica EZ Viewing Tube – Integrated Eyepieces (Continued)

 You can now safely rotate the Leica EZ viewing tube by loosening the captive thumbscrew, rotating the viewing tube, and tightening the captive thumbscrew again.

 The eyepieces are integrated into the Leica EZ viewing tube and preset; therefore, there is no need to adjust or install the eyepieces.

Continue with the "[Eyecups](#)" section on [page 25](#).



Leica Standard Viewing Tube – Separated Eyepieces



Standard viewing tubes; tubes do not include eyepieces yet

i The standard viewing tube has a rotatable dovetail. Therefore, you can now rotate the standard viewing tube freely in any orientation.

1. Insert the eyepieces into the tubes.




2. Secure the eyepieces in the tubes by tightening the silver screws at the bottom of the tubes using a standard Phillips screwdriver (not included in the delivery package).




i The eyepieces will still rotate, but they will be captive in the eyetubes.



Eyecups

 If you wear eyeglasses for microscope viewing, keep the rubber eyecups folded down. If you do not wear eyeglasses, you may find it useful to unfold the rubber eyecups in order to help block out ambient room light.



 If you have purchased a standard microscope configuration, you will notice that the objectives are already installed on the nose-piece and the substage condenser is already installed on the stand. In this case, go to section "[Operation](#)" on [page 30](#). If you purchased your Leica DM750 in the form of individual components rather than the standard configuration, please continue with the "[Installing Objectives](#)" section on [page 26](#).



Installing Objectives

Installing objectives




When rotating the objective nosepiece, always use the knurled ring on the objective nosepiece.

As you rotate the nosepiece clockwise, attach the objectives by screwing them into the nosepiece holes starting with the lowest magnification and advancing to the highest magnification.



Installing the Substage Condenser

Substage condenser

 The Leica DM750 has an open substage condenser mount, therefore the condenser needs to be installed.




The substage condenser mount is open

1. Move the specimen stage upwards as far as it will go by using the coarse focusing knob on the side of the microscope.
2. Move the condenser holder into the lowest position by using the condenser focusing knob on the left side of the stage mount.
3. Unscrew the two setscrews (or wingscrews for Koehler stands) on the condenser holder.



2. Move the condenser holder into the lowest position by using the condenser focusing knob on the left side of the stage mount.




Installing the Substage Condenser (Continued)

4. Push the substage condenser under the specimen stage into the holder by aligning the locating pin on the bottom of the condenser with the slot on the rear side of the fork.




5. Move the condenser into the highest position by using the condenser focusing knob on the left side of the stage mount.



6. Tighten the two setscrews using the tool included in the delivery package (or, for a Koehler stand, tighten the two wingscrews) so that the upper lens of the condenser is centered under the objective in working position and the substage condenser is thus roughly centered.



 You will center the condenser more accurately when you get to the "[Complete Condenser Centering](#)" section on [page 31](#).

Set!



Turning on the microscope

Work surface



Always use your microscope on a hard, stable surface.

Power cord



If the power cord is not already attached, attach it securely to the back of the microscope.



USB power connector

The Leica DM750 has a 5 V/1.5 A USB power connector in the center of the cord wrap. This can be used to power some Leica cameras or other devices requiring 5 V/1.5 A.

Setting the Illumination Intensity



Set the illumination to the lowest setting to start with, using the controller on the bottom left of the stand. The illumination control knob allows you to adjust the intensity of light produced by the illumination system.




Plug in and turn on the microscope

1. Plug the power cable of the microscope into a corresponding grounded socket. A grounded 3-wire cord is provided.
2. Switch on the microscope using the switch at the bottom right of the microscope stand.




Complete Condenser Centering



 If you have purchased a standard Leica DM750 configuration, the condenser has already been precentered by Leica Microsystems.

1. Open the condenser aperture by rotating the knurled ring on the condenser to the right.




 Make sure that the condenser is in the highest position.

2. Write an "X" on a piece of paper in the size of a business card and place it onto the light output of the microscope stand in such a way that the "X" is centered over the illumination.



Not centered



Centered

3. Look at the X through the eyepieces and center it in the field of view by tightening the setscrews with the tool included in the delivery package (or, for a Leica DM750 with Koehler illumination, tighten the knurled screws).

Using the Condenser

Using the condenser



The condenser is furnished with an iris diaphragm, which can be adjusted to match the effective numerical aperture of the objective

1. To open and close this diaphragm, simply turn the knurled ring on the condenser to the right or left so that the line on the ring is aligned with the objective magnification used.
2. Open the iris diaphragm of the condenser completely at first by turning the condenser ring all the way to the right.



Match the line of the rotating ring with the objective magnification in use.



Prepare to View a Specimen Slide

1. Position a specimen slide on the specimen stage by sliding it under the slide grips.
2. Using the X/Y-stage control, position the specimen slide such that a part of the specimen is under the objective used.



Slide grips hold the slide in place.





Slide grips




Focusing

1. Rotate the objective nosepiece (using the knurled ring) in such a way that the objective with the lowest magnification level is rotated into the working position.
2. Move the specimen stage upwards by turning the coarse focusing knob as far as it will go to the maximum position.



Coarse focus adjustment knob

3. Look into the eyepieces and adjust the illumination intensity to a level that is comfortable for your eyes.

 The stand of the Leica DM750 has been calibrated at the factory so that the focus can be found from this position within 1.5 rotations of the fine focus.

4. Bring the specimen into focus using the fine focusing knob.



Viewing Tube Adjustment

Adjust the eyetubes

1. Adjust the tubes to your interpupillary distance. Fold or unfold the eyetubes to decrease or increase the distance between the eyepieces until you see one illuminated circle.



i The viewing tubes maintain a constant tube length for all interpupillary settings. This means that a change of interpupillary distance does not affect parfocality, magnification, or calibrations that depend on magnification.

Leica EZ viewing tube

- If you are using a Leica EZ viewing tube, which has the eyepieces integrated with the eyetubes, no additional adjustments are necessary. Be sure you are wearing your eyeglasses or contact lenses.
- If you have a Leica DM750 with standard illumination (without Koehler field diaphragm), continue with the "[Oil Immersion Technique](#)" section on [page 39](#).
- If you have a Leica DM750 with standard illumination (without Koehler field diaphragm), continue with the "[Oil Immersion Technique](#)" section on [page 39](#).
- If you have a Leica DM750 with Koehler field diaphragm, continue with the "[Koehler Configuration](#)" section on [page 37](#).

Standard viewing tube with two fixed eyepieces

- If you are using a standard viewing tube with two fixed eyepieces (no focusing eyepieces), no additional adjustments necessary. Be sure you are wearing your eyeglasses or contact lenses.

- If you have a Leica DM750 with standard illumination (without Koehler field diaphragm), continue with the "[Oil Immersion Technique](#)" section on [page 39](#).
- If you have a Leica DM750 with Koehler field diaphragm, continue with the "[Koehler Configuration](#)" section on [page 37](#).


Standard viewing tube with one or two focusing eyepieces

If you are using a standard viewing tube with one or two focusing eyepieces, you need to make some adjustments.

1. Set the focusing eyepieces to "0".



Viewing Tube Adjustment (Continued)

 If you are comfortable wearing your corrective lenses (contact lenses or eyeglasses) for microscope viewing, leave them on and your adjustments will be minimal.

- Using the fine focus adjusting knob, focus on the specimen while looking through only one of the eyepieces (when using one eyepiece that is focusable and one that is not focusable, look through the one that is not focusable). To help focus, cover or close the other eye.



- Now look with the other eye just through the other eyepiece (focusing eyepiece). This time, focus the specimen by using the focusing capability in the focusing eyepiece.



When doing so, do not change the height of the specimen stage.

- Grip the knurled ring on the focusing eyepiece with one hand and rotate the top of the eyepiece with the other hand until the specimen is in focus for this eye and

this focusing eyepiece. This corrects for any vision differences between your right eye and left eye.

- Now, switch to an objective with a high magnification level (no oil objective) and bring the microscope into focus while looking through the eyepieces with both eyes.



The higher magnifications have a shallower depth of field. Therefore, after focusing with a high magnification, you will find that when you change to lower magnifications, you only have to adjust the fine focus slightly, if at all.

- If you use a DM750 with standard illumination, continue with the "[Oil Immersion Technique](#)" section on [page 39](#).
- If your DM750 has a Koehler configuration, please continue with the next section, "[Koehler Configuration](#)", on [page 37](#).



Koehler Configuration

i If your Leica DM750 is fitted with a field diaphragm for Koehler Illumination, use the following procedure to assure condenser centering and focus.

1. Connect the Koehler field diaphragm to the base of the microscope so that the diaphragm leaves are in the field of view when you look through the eyepieces.




Close the Koehler field diaphragm




Close the Koehler field diaphragm

2. Focus the leaves of the field diaphragm using the condenser focusing knob on the left side of the stage mount.




Condenser focusing knob






Leaves are in sharp focus

Koehler Configuration (Continued)

3. Turn the condenser centering wingscrews simultaneously to center the image of the field diaphragm.



4. Open the field diaphragm until the diaphragm leaves are just outside of the field of view.



Oil Immersion Technique

1. Search for the area on the specimen slide that you want to study.



2. Move the specimen stage down to the lowest position using the coarse adjustment knob.

3. Add a drop of immersion oil from Leica to the area of the specimen slide you are studying.



4. Rotate the oil immersion objective (the objective labeled "OIL") to the working position.



Oil Immersion Technique (Continued)

5. Slowly move the specimen stage upwards using the coarse adjustment knob until the oil drop on the slide comes into direct contact with the lens of the oil immersion objective.



6. Hold the knurled ring on the objective turret and swing the objective back and forth to eliminate air bubbles. Then bring the oil objective into the final position, so that the oil drop is between the front lens of the objective and the specimen slide.

7. Look through the microscope and slowly rotate the fine adjustment knob, so that the specimen stage moves upwards until the specimen is in focus.
8. If you are finished working with the oil immersion objective, clean the front side of the objective, the specimen slide, and all other surfaces that have come into contact with the oil, following the instructions in the "Care of the Microscope" section on page 44.

Time Delay Shutoff

i The Leica DM750 is equipped with a time delay shutoff capability, which automatically turns the illumination off after 2 hours of no changes in the intensity control.

- All four-position nosepiece stands have the time delay shutoff enabled as a default (mostly educational applications).
- All five-position nosepiece stands have the time delay shutoff disabled as a default (mostly clinical applications).

Changing the status of the time delay shutoff

1. Rotate the intensity regulator to the lowest level.
2. Switch on the instrument.
3. Rotate the intensity regulator to the highest level and then back to the lowest level within five seconds.



- The LED Illumination will flash to indicate the time delay shutoff status was changed.
- The LED Illumination will flash two times slowly then stay on when you disable the Time Delay Shutoff.
- The LED Illumination will flash three times quickly then stay on when you enable the Time Delay Shutoff.

i When you turn the power off and then turn the power on again, the system will be in the last Time Delay Shutoff status (Enabled or Disabled) and you will not see any blinking.

Go!



Ready! Set! Go!

Now all you need to do is change objectives, set the condenser aperture (and field diaphragm if you have a DM750 Koehler Stand) appropriately for the objective magnification you are using, and enjoy the view!

Care of the Microscope

General Maintenance

General



Always carry the microscope using two hands. There is a handle on the back of the microscope and an undercut in the front for this purpose.



- The cord wrap allows you to wrap the cord in such a way that only the length you need is extended.



- If any optical surface becomes coated with dust or dirt, clean the surface by blowing it off with a syringe or brushing it off with a camel hair brush before attempting to wipe the surface clean.
- Optical surfaces should be cleaned with a lint-free cloth, lens tissue, or cotton swab moistened with a commercially available glass cleaner.
- It is very important to avoid the excessive use of solvents, so use them sparingly. The lint-free cloth, lens tissue or cotton swab should be moistened with solvent, but not be wet enough for the solvent to seep around the lens.
- Keep all optical components clean. Cleanliness is important for maintaining good optical performance.
- The microscope should always be covered with the plastic dust cover (provided with the instrument) when it is not in use.



General Maintenance (Continued)

- No part of the microscope is quite so vulnerable to collecting dirt, dust, and oil as the front lens of the objective. Whenever you encounter lack of contrast, cloudiness or poor definition, carefully check the condition of the front lens with a magnifier.
- Cleaning 40× and 100× objectives requires more care. Note: To achieve the high degree of flatness obtained with higher magnification objectives, the objective has a small concave front lens of fairly short radius or curvature. The surface of this front lens can be readily cleaned with a toothpick covered with a cotton tip, or with a small cotton swab. Moisten the cotton with commercially available glass cleaner. Wipe the front lens lightly without applying undue force or scrubbing action. Make sure that the cotton tip contacts the concave lens surface. Check the objective with a magnifier after cleaning.
- If you need to remove the microscope's viewing body, be careful not to accidentally touch the outer lens surface (located on the underside of the body). Fingerprints on this surface will reduce image clarity. This lens can be cleaned in the same manner as objectives and eyepieces.

Illumination

- The Leica DM750 uses LED illumination. Therefore, no lamp changing is required for the life of the microscope.



Troubleshooting

Troubleshooting

Stand	
The microscope does not respond.	Make sure that the socket has voltage.
	Check the cable connections.
	Make sure that the stand is connected to the power supply.
	Check whether the fuse is defective and replace it if necessary (see page 15).
Focus	
The specimen cannot be brought into focus.	Use the correct immersion medium.
	Lay the specimen with the cover glass toward the top.
	Make sure that the cover slip thickness is correct and that it meets the specifications on the objective.
Dark field	
No definite DF contrast is possible.	Make sure that a DF objective is being used.
	The objective aperture is too high (maximum 0.75/ 1.10); if necessary, reduce the objective aperture through the iris diaphragm on the objective.
	Check the condenser centering.
	Open the aperture diaphragm completely.
The image is not uniformly illuminated.	The objective magnification is too weak. Use a higher magnification.
Unwanted stray light.	Clean the specimen and neighboring lens surfaces.





Troubleshooting (Continued)



Polarization	
Polarization contrast cannot be adjusted.	Bring the polarizer and analyzer into cross position until they reach maximum darkness (without specimen).
Phase contrast	
Phase contrast cannot be adjusted.	The specimen is too thick, too thin, or too brightly stained.
	Refractive indexes of the mounting medium and specimen are identical, so that there is no phase jump.
	The cover glass is not placed uniformly.
	Check that the correct light ring is positioned.
	Check the centering of the light rings.
	Check the condenser centering.
	Open the aperture diaphragm completely.
Specimen stage	
Positioning range of the stage in the x-direction decreases after working a long time.	Move the mechanical stage with coaxial drive all the way to the left.
	Manually press the screw that holds the mechanical stage even farther to the left, as far as it will go. Then move the mechanical stage with coaxial drive all the way to the right.
	Manually press the screw that holds the mechanical stage even farther to the right, as far as it will go.





Dimensions









The statement by Ernst Leitz in 1907, *"With the User, For the User,"* describes the fruitful collaboration with end users and driving force of innovation at Leica Microsystems. We have developed five brand values to live up to this tradition: Pioneering, High-end Quality, Team Spirit, Dedication to Science, and Continuous Improvement. For us, living up to these values means: **Living up to Life.**

INDUSTRY DIVISION

The Leica Microsystems Industry Division's focus is to support customers' pursuit of the highest quality end result. Leica Microsystems provide the best and most innovative imaging systems to see, measure, and analyze the microstructures in routine and research industrial applications, materials science, quality control, forensic science investigation, and educational applications.



13IDE13060EN_1 · Copyright © by Leica Microsystems (Schweiz) AG, CH-9435 Heerbrugg, 2015. Subject to modifications. LEICA and the Leica Logo are registered trademarks of Leica Microsystems IR GmbH.